

**Certificate of Analysis
GM501 VitriStore Freeze**

Batchnumber	VITRI-1901
Production date	14 January 2019
Expiry date	31 January 2020
Batchnumber HSA	B3AKC00053
Total volume (Kits)	250

GM501 VitriStore Pre-

Vitrification	Results	Specifications
pH	7,26	7,20-7,40
Sterility test	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	281	270-290 mOsm/kg

GM501 VitriStore Freeze 1

Results	Specifications
pH	7,26
Sterility test	PASSED
Endotoxin	< 0,1

GM501 VitriStore Freeze 2

Results	Specifications
pH	7,28
Sterility test	PASSED
Endotoxin	< 0,2

Mouse Embryo Assay (1-cell)	100	≥ 80 %
-----------------------------	-----	--------

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	29 January 2019
--------------	-----------------

Released by	Dr. Julia Heinzmann
-------------	---------------------

Signature

p.p. Julia Heinzmann