

## Certificate of Analysis GM501 VitriStore Thaw

|                     |             |
|---------------------|-------------|
| Batchnumber         | VITRI-1906  |
| Production date     | 02 May 2019 |
| Expiry date         | 31 May 2020 |
| Batchnumber HSA     | B3ALC00323  |
| Total volume (Kits) | 200         |

| GM501 VitriStore Thaw 1 | Results | Specifications  |
|-------------------------|---------|-----------------|
| pH                      | 7,3     | 7,20-7,40       |
| Sterility test          | PASSED  | PASSED          |
| Endotoxin               | < 0,1   | < 0,25 EU/mL    |
| Osmolality              | 832     | 805-850 mOsm/kg |

| GM501 VitriStore Thaw 2 | Results | Specifications  |
|-------------------------|---------|-----------------|
| pH                      | 7,34    | 7,20-7,40       |
| Sterility test          | PASSED  | PASSED          |
| Endotoxin               | < 0,117 | < 0,25 EU/mL    |
| Osmolality              | 553     | 535-565 mOsm/kg |

| GM501 VitriStore Thaw 3 | Results | Specifications  |
|-------------------------|---------|-----------------|
| pH                      | 7,33    | 7,20-7,40       |
| Sterility test          | PASSED  | PASSED          |
| Endotoxin               | < 0,112 | < 0,25 EU/mL    |
| Osmolality              | 418     | 405-435 mOsm/kg |

| GM501 VitriStore Thaw 4 | Results | Specifications  |
|-------------------------|---------|-----------------|
| pH                      | 7,33    | 7,20-7,40       |
| Sterility test          | PASSED  | PASSED          |
| Endotoxin               | < 0,1   | < 0,25 EU/mL    |
| Osmolality              | 282     | 270-290 mOsm/kg |

|                             |       |        |
|-----------------------------|-------|--------|
| Mouse Embryo Assay (1-cell) | 90,48 | ≥ 80 % |
|-----------------------------|-------|--------|

*Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.*

|              |             |
|--------------|-------------|
| Release date | 20 May 2019 |
|--------------|-------------|

|             |                     |
|-------------|---------------------|
| Released by | Dr. Julia Heinzmann |
|-------------|---------------------|

Signature

*p.p. Julia Heinzmann*