

Certificate of Analysis GM501 VitriStore Freeze

Batchnumber	VITRI-2109
Production date	25 May 2021
Expiry date	31 May 2022
Batchnumber HSA	B3ALE00033
Total volume (Kits)	120

GM501 VitriStore Pre-

Vitrification	Results	Specifications
pH	7,24	7,20-7,40
Sterility test	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	282	270-290 mOsm/kg

GM501 VitriStore Freeze 1

Vitrification	Results	Specifications
pH	7,24	7,20-7,40
Sterility test	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL

GM501 VitriStore Freeze 2

Vitrification	Results	Specifications
pH	7,24	7,20-7,40
Sterility test	PASSED	PASSED
Endotoxin	< 0,2	< 0,25 EU/mL

Mouse Embryo Assay (1-cell)	100	≥ 80 %
-----------------------------	-----	--------

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	14 June 2021
--------------	--------------

Released by	Dr. Julia Heinzmann
-------------	---------------------

Signature

p.p. Julia Heinzmann