

Certificate of Analysis GM501 VitriStore Freeze

Batchnumber	VITRI-2415
Production date	23 August 2024
Expiry date	31 August 2025
Batchnumber HSA	30
Total volume (Kits)	B03H100413

Parameter	Results	Specifications
GM501 VitriStore Pre-Vitrification		
pH	7,24	7,20-7,40
Sterility test ¹	No growth	No growth
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	283	270-290 mOsm/kg
Concentration HSA	20,3	16,0-24,0 g/l
GM501 VitriStore Freeze 1		
pH	7,28	7,20-7,40
Sterility test ¹	No growth	No growth
Endotoxin	< 0,1	< 0,25 EU/mL
Concentration HSA	16,4	12,8-19,2 g/l
GM501 VitriStore Freeze 2		
pH	7,21	7,20-7,40
Sterility test ¹	No growth	No growth
Endotoxin	< 0,2	< 0,25 EU/mL
Concentration HSA	10,8	8,3-12,4 g/l
Mouse Embryo Assay (1-cell)	95	≥ 80 %


Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

¹ (according to the current Ph.Eur.2.6.1 / USP <71>)

Release date 09 September 2024

Released by Nicole Geiger

Signature



REV 06_00 / 2023-09-15