

Certificate of Analysis GM501 VitriStore Thaw

Batchnumber	VITRI-2518
Production date	18 June 2025
Expiry date	30 June 2026
Total batch size	150 Kits
Batchnumber HSA	B03H121903

Parameter	Results	Specifications
GM501 VitriStore Thaw 1		
pH	7,31	7,20-7,40
Sterility test ¹	No growth	No growth
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	837	805-850 mOsm/kg
Concentration HSA	17,7	14,5-21,7 g/l
GM501 VitriStore Thaw 2		
pH	7,32	7,20-7,40
Sterility test ¹	No growth	No growth
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	556	535-565 mOsm/kg
Concentration HSA	19,2	15,2-22,8 g/l
GM501 VitriStore Thaw 3		
pH	7,28	7,20-7,40
Sterility test ¹	No growth	No growth
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	416	405-435 mOsm/kg
Concentration HSA	19,7	15,5-23,3 g/l
GM501 VitriStore Thaw 4		
pH	7,26	7,20-7,40
Sterility test ¹	No growth	No growth
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	283	270-290 mOsm/kg
Concentration HSA	20	16,0-24,0 g/l

Mouse Embryo Assay (1-cell)	90	≥ 80 %
-----------------------------	----	--------

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

¹(according to the current Ph.Eur.2.6.1 / USP <71>)

Release date	16 July 2025
--------------	--------------

Released by	Joana Sturm
-------------	-------------

Signature



REV 06_00 / 2023-09-15