

**Certificate of Analysis
GM501 VitriStore Thaw**

Batchnumber	VITRI-2602
Production date	21 January 2026
Expiry date	31 January 2027
Total batch size	195 Kits
Batchnumber HSA	B03J103333

Parameter	Results	Specifications
GM501 VitriStore Thaw 1		
pH	7,25	7,20-7,40
Sterility test ¹	No growth	No growth
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	826	805-850 mOsm/kg
Concentration HSA	16,7	14,5-21,7 g/l
GM501 VitriStore Thaw 2		
pH	7,24	7,20-7,40
Sterility test ¹	No growth	No growth
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	550	535-565 mOsm/kg
Concentration HSA	17,4	15,2-22,8 g/l
GM501 VitriStore Thaw 3		
pH	7,21	7,20-7,40
Sterility test ¹	No growth	No growth
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	417	405-435 mOsm/kg
Concentration HSA	17,9	15,5-23,3 g/l
GM501 VitriStore Thaw 4		
pH	7,25	7,20-7,40
Sterility test ¹	No growth	No growth
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	280	270-290 mOsm/kg
Concentration HSA	18,6	16,0-24,0 g/l

Mouse Embryo Assay (1-cell) 95 ≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

¹(according to the current Ph.Eur.2.6.1 / USP <71>)

Release date	18 February 2026
Released by	Dr. René Richarz
Signature	

REV 06_00 / 2023-09-15