

Certificate of Analysis GM501 VitriStore Freeze

Batchnumber	VITRI-2608
Production date	08 May 2026
Expiry date	31 May 2027
Batchnumber HSA	B03J125033
Total volume (Kits)	100

Product was sterilized by aseptic processing techniques

Parameter	Results	Specifications
GM501 VitriStore Pre-Vitrification		
pH	7,29	7,20-7,40
Sterility test ¹	No growth	No growth
Endotoxin	< 0,15	< 0,25 EU/mL
Osmolality	284	270-290 mOsm/kg
Concentration HSA	19,9	16,0-24,0 g/l
GM501 VitriStore Freeze 1		
pH	7,28	7,20-7,40
Sterility test ¹	No growth	No growth
Endotoxin	< 0,15	< 0,25 EU/mL
Concentration HSA	16,1	12,8-19,2 g/l
GM501 VitriStore Freeze 2		
pH	7,25	7,20-7,40
Sterility test ¹	No growth	No growth
Endotoxin	< 0,15	< 0,25 EU/mL
Concentration HSA	10,6	8,3-12,4 g/l
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

¹ (according to the current Ph.Eur.2.6.1 / USP <71>)

Release date 29 May 2026

Released by Dr. René Richarz

Signature



REV 06_00 / 2023-09-15