

**Certificate of Analysis  
GM501 WASH**

Batchnumber	WASH-1703
Production date	13 March 2017
Expiry date	30 September 2017
Total volume (mL)	25500
Batchnumber HSA	A4DF500171

	<b>Results</b>	<b>Specifications</b>
pH	7,27	7,20-7,50
Osmolality	280	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95	≥ 80 %

*Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.*

Release date 29 March 2017

Released by Dr. Julia Heinzmann

Signature

*p.p. Julia Heinzmann*