

**Certificate of Analysis
GM501 WASH**

Batchnumber	WASH-1901
Production date	11 January 2019
Expiry date	31 July 2019
Batchnumber HSA	B3AKC00053
Total volume (mL)	22400

	Results	Specifications
pH	7,21	7,20-7,50
Osmolality	279	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 28 January 2019

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann