

**Certificate of Analysis
GM501 WASH**

Batchnumber	WASH-1911
Production date	08 August 2019
Expiry date	29 February 2020
Batchnumber HSA	B3ALC00833
Total volume (mL)	16300

	Results	Specifications
pH	7,3	7,20-7,50
Osmolality	284	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	23 August 2019
--------------	----------------

Released by	Dr. Julia Heinzmann
-------------	---------------------

Signature

p.p. Julia Heinzmann