

**Certificate of Analysis  
GM501 SpermActive**

Batchnumber	WASH-2004PRG
Production date	01 April 2020
Expiry date	31 October 2020
Batchnumber HSA	B3ALD01343
Total volume (mL)	40500

	<b>Results</b>	<b>Specifications</b>
pH	7,23	7,20-7,50
Osmolality	279	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

*Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.*

Release date 16 April 2020

Released by Vanessa Bierlich

Signature