

Certificate of Analysis
GM501 WASH with Phenol Red and Gentamicin

Batchnumber	WASH-2104PRG
Production date	12 March 2021
Expiry date	30 September 2021
Batchnumber HSA	B3ALE00033
Total volume (mL)	18500

	Results	Specifications
pH	7,25	7,20-7,50
Osmolality	277	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 29 March 2021

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann