

Certificate of Analysis

GM501 WASH with Phenol Red and Gentamicin

Batchnumber	WASH-2111PRG
Production date	04 October 2021
Expiry date	30 April 2022
Batchnumber HSA	B3ALE00033
Total volume (mL)	31400

	Results	Specifications
pH	7,27	7,20-7,50
Osmolality	280	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	90,48	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	19 October 2021
--------------	-----------------

Released by	Dr. Julia Heinzmann
-------------	---------------------

Signature

p.p. Julia Heinzmann