

Certificate of Analysis GM501 WASH

Batchnumber	WASH-2206
Production date	27 June 2022
Expiry date	31 December 2022
Batchnumber HSA	B3ALE00033
Total volume (mL)	11300

	Results	Specifications
pH	7,31	7,20-7,50
Osmolality	283	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	12 July 2022
--------------	--------------

Released by	Dr. Julia Heinzmann
-------------	---------------------

Signature

p.p. Julia Heinzmann