

**Certificate of Analysis
GM501 WASH**

Batchnumber WASH-2208
 Production date 26 September 2022
 Expiry date 31 March 2023
 Batchnumber HSA B03G002753
 Total volume (mL)

	Results	Specifications
pH	7,28	7,20-7,50
Osmolality	281	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	90,48	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 12 October 2022
 Released by Vanessa Bierlich

Signature