

**Certificate of Analysis
GM501 WASH**

Batchnumber	WASH-2303
Production date	07 February 2023
Expiry date	31 August 2023
Batchnumber HSA	B03G002753
Total volume (mL)	12600

	Results	Specifications
pH	7,2	7,20-7,50
Osmolality	280	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,09	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	90	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 23 February 2023

Released by Nicole Geiger

Signature

