

**Certificate of Analysis
GM501 WASH**

Batchnumber	WASH-2312
Production date	15 November 2023
Expiry date	31 May 2024
Batchnumber HSA	B03H025323
Total volume (mL)	11100

Parameter	Results	Specifications
pH	7,39	7,20-7,50
Osmolality	278	270-290 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	95,24	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	30 November 2023
Released by	Vanessa Bierlich
Signature	

P.P. 

REV 06_00 / 2023-09-15