

**Certificate of Analysis  
GM501 SpermActive**

Batchnumber	WASH-2313PRG
Production date	20 December 2023
Expiry date	30 June 2024
Batchnumber HSA	B03H025323
Total volume (mL)	29600

Parameter	Results	Specifications
pH	7,36	7,20-7,50
Osmolality	278	270-290 mOsm/kg
Bacterial screen	No growth	PASSED
Fungal screen	No growth	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

*Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.*

Release date	08 January 2024
Released by	Nicole Geiger
Signature	

pp. N. Geiger

REV 06\_00 / 2023-09-15