

Certificate of Analysis

GM501 WASH with Phenol Red and Gentamicin

Batchnumber	WASH-2402PRG
Production date	13 March 2024
Expiry date	30 September 2024
Batchnumber HSA	B03H025323
Total volume (mL)	37400

Parameter	Results	Specifications
pH	7,33	7,20-7,50
Osmolality	277	270-290 mOsm/kg
Bacterial screen	NO GROWTH	No growth
Fungal screen	NO GROWTH	No growth
Endotoxin	< 0,05	< 0,25 EU/mL
HSA concentration	4,7	4,0-6,0 mg/ml
Mouse Embryo Assay (1-cell)	95,24	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	28 March 2024
Released by	Nicole Geiger
Signature	