

**Certificate of Analysis  
GM501 SpermActive**


Batchnumber WASH-2404PRG  
 Production date 10 June 2024  
 Expiry date 31 December 2024  
 Batchnumber HSA B03H100413  
 Total volume (mL) 24600

Parameter	Results	Specifications
pH	7,33	7,20-7,50
Osmolality	280	270-290 mOsm/kg
Bacterial screen	No growth	No growth
Fungal screen	No growth	No growth
Endotoxin	< 0,05	< 0,25 EU/mL
HSA concentration	4,6	4,0-6,0 mg/ml
Mouse Embryo Assay (1-cell)	100	≥ 80 %

*Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.*

Release date 25 June 2024  
 Released by Nicole Geiger

Signature



REV 07\_00 / 2024-02-08