


**Certificate of Analysis  
GM501 SpermActive**

Batchnumber WASH-2406PRG  
 Production date 02 September 2024  
 Expiry date 31 March 2025  
 Batchnumber HSA B03H100413  
 Total volume (mL) 27600

Parameter	Results	Specifications
pH	7,26	7,20-7,50
Osmolality	281	270-290 mOsm/kg
Endotoxin level	< 0,05	< 0,25 EU/ml
Concentration HSA	4,7	4,0-6,0 g/l
Concentration Gentamicin	12	5-15 mg/l
Sterility test (performed according to the current Ph.Eur.2.6.1 / USP <71>)	No growth	No growth
Mouse Embryo Assay (1-cell to expanded blastocyst within 96 hours)	95	≥ 80 %

*Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.*

Release date 18 September 2024  
 Released by Nicole Geiger  
 Signature



REV 08\_00 / 2024-09-17