

**Certificate of Analysis  
GM501 WASH**

Batchnumber WASH-2410  
 Production date 02 December 2024  
 Expiry date 30 June 2025  
 Total volume (ml) 6000  
 Batchnumber HSA B03H100413

Product was sterilized by aseptic processing techniques

Parameter	Results	Specifications
pH	7,29	7,20-7,50
Osmolality	281	270-290 mOsm/kg
Endotoxin level	< 0,05	< 0,25 EU/ml
Concentration HSA	4,7	4,0-6,0 g/l
<b>Sterility test</b> (performed according to the current Ph.Eur.2.6.1 / USP <71>)	No growth	No growth
<b>Mouse Embryo Assay</b> (1-cell to expanded blastocyst within 96 hours)	95,24	≥ 80 %

*Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.*

Release date 18 December 2024

Released by Nicole Geiger

Signature

