

## Certificate of Analysis

### GM501 WASH with Phenol Red and Gentamicin

Batchnumber	WASH-2503PRG
Production date	09 April 2025
Expiry date	31 October 2025
Total volume (mL)	36500
Batchnumber HSA	B03H121903
Product was sterilized by aseptic processing techniques	

Parameter	Results	Specifications
pH (@37°C - 6% CO <sub>2</sub> )	7,28	7,20-7,50
Osmolality	281	270-290 mOsm/kg
Endotoxin level	< 0,05	< 0,25 EU/ml
Concentration HSA	4,7	4,0-6,0 g/l
Concentration gentamicin	12	5-15 mg/l
Sterility test (performed according to the current Ph.Eur.2.6.1 / USP <71>)	No growth	No growth
Mouse Embryo Assay (1-cell to expanded blastocyst within 96 hours)	100	≥ 80 %

*Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.*

Release date	24 April 2025
Released by	Joana Sturm
Signature	



REV 09\_00 / 2025-03-19