

**Certificate of Analysis
GM501 SpermActive**

Batchnumber **WASH-2604PRG**
 Production date 22 April 2026
 Expiry date 31 October 2026
 Total volume (mL) 28900
 Batchnumber HSA B03J125033

Product was sterilized by aseptic processing techniques ☒

Parameter	Results	Specifications
pH (@37°C - 6% CO ₂)	7,29	7,20-7,50
Osmolality	282	270-290 mOsm/kg
Endotoxin level	< 0,05	< 0,25 EU/ml
Concentration HSA	4,8	4,0-6,0 g/l
Concentration gentamicin	9	5-15 mg/l
Sterility test (performed according to the current Ph.Eur.2.6.1 / USP <71>)	No growth	No growth
Mouse Embryo Assay (1-cell to expanded blastocyst within 96 hours)	95	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 08 May 2026
 Released by Nicole Geiger
 Signature 