


Certificate of Analysis
GM501 WASH with Phenol Red and Gentamicin

Batchnumber	WASH-2607PRG
Production date	17 August 2026
Expiry date	28 February 2027
Total volume (mL)	18200
Batchnumber HSA	B03J125033
Product was sterilized by aseptic processing techniques	

Parameter	Results	Specifications
pH (@37°C - 6% CO ₂)	7,31	7,20-7,50
Osmolality	282	270-290 mOsm/kg
Endotoxin level	< 0,05	< 0,25 EU/ml
Concentration HSA	4,7	4,0-6,0 g/l
Concentration gentamicin	11	5-15 mg/l
Sterility test (performed according to the current Ph.Eur.2.6.1 / USP <71>)	No growth	No growth
Mouse Embryo Assay (1-cell to expanded blastocyst within 96 hours)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	02 September 2026
Released by	Nicole Geiger
Signature	

REV 09_00 / 2025-03-19