

Mouse Embryo Assay Certificate of Analysis

Manufacturer		Microtech
Product		Micropipette
Batchnumber		2003041
Expiry date		03/2023
Mouse Embryo Test	<u>Result</u>	<u>Specifications</u>
	100	≥ 80 %

Assay system requested by customer:

1mL of culture medium was placed in a tube with the test article for 30-minutes at 37°C and 5 % CO₂. Post incubation three 12.5 µl drops of the culture medium was extracted from the test article tube and placed in the corresponding wells of a culture dish; 7 one cell mouse embryos were added to each of the three wells and cultured for 96-hours.

Control assay method and results:

21 one cell (B₆C₃F₁ X B₆D₂F₁) embryos were cultured in triplicate micro drops of culture

21/21 (100%)

1-cell to 2-cell within 24 hr

21/21 (100%)

1-cell to expanded blastocyst within 96 hr

For a valid assay, Embryotech™ requires at least 70 % of one cell stage control embryos to develop to expanded blastocyst within 96-hours.

Test assay method and results:

21 one cell (B₆C₃F₁ X B₆D₂F₁) embryos were cultured in triplicate micro drops of culture medium that was extracted from the test article:

21/21 (100%)

1-cell to 2-cell within 24 hr

21/21 (100%)

1-cell to expanded blastocyst within 96 hr

Summary of observations:

All test and control embryos were selected randomly from a common pool of freshly collected embryos and were cultured in the same incubator at 37°C and 5% CO₂. 100 percent of the control embryos developed to the expanded blastocyst stage within 96-hours. 100 percent of the test embryos cultured in the extracted culture medium developed to the expanded blastocyst stage within 96-hours.

Release date

31/03/2020

Released by

Peggy Kreyser

Signature



Rev01_00 / 2019-04-15