

**Certificate of Analysis  
GM508 CultActive**

Batchnumber	CACT-1707
Production date	08 November 2017
Expiry date	31 May 2018
Total volume (mL)	990

	<b>Results</b>	<b>Specifications</b>
pH	7,35	7,00-7,60
Osmolality	420	400-450 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Release date	23 November 2017
Released by	Dr. Julia Heinzmann
Signature	

*p.p. Julia Heinzmann*