

**Certificate of Analysis
GM501 Gradient 100%**

Batchnumber	GSS-1703
Production date	12 July 2017
Expiry date	31 January 2019
Total volume (mL)	9250

	Results	Specifications
pH	7,59	7,20-7,60
Osmolality	306	300-330 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,495	< 0,50 EU/mL
Density	1,1195	1,1150-1,1250 g/mL
Viscosity in cP (26°C)	1,56	< 1,75 cP
Sperm survival Test	97	≥ 80 %

Release date 27 July 2017

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann