

Certificate of Analysis GM501 PVP

Batchnumber	PVP-1721
Production date	27 November 2017
Expiry date	31 August 2018
Total volume (kits)	100
Batchnumber HSA	B3AKA00593

	Results	Specifications
pH	7,45	7,20-7,60
Osmolality	294	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 1,00 EU/mL
Viscosity in cP (26°C)	200,4	> 170 cP
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	12 December 2017
Released by	Dr. Julia Heinzmann
Signature	

p.p. Julia Heinzmann