




**SparMED Aps  
Ryttermarken 2  
3520 Farum  
Denmark**



140 Hale Street  
Haverhill, MA 01830  
[qc@embryotech.com](mailto:qc@embryotech.com)

**ELI Accession Number:** S2911-0317SPAR

**Date of completion:** 03-24-2017

**Lot number:** 07730

**Order numbers:** OOPW-SW02, OOPW-CW05, OOPW-OT10

**Description of test article(s):** Oosafe® 6 Well Dish with Straw Holder, Center Well Dish, OPU Tube

**Assay system requested by customer:** Prepared sperm was added to the test articles (3 test articles pooled) and incubated for 24-hours. Post incubation the sperm wash medium with sperm was extracted from the test articles and pooled. The forward progressive motility was read and recorded at 24-hours.

**Results:**

Test method: SOP/TSG/ELI/008	Specification	Initial	Result % 24hr	SMI Value	Pass/Fail
---------------------------------	---------------	---------	------------------	--------------	-----------

Test Article	Specification	Initial	Result % 24hr	SMI Value	Pass/Fail
Test Article	SMI $\geq$ 0.75	88%	85%	0.97	Pass
Control	$\geq$ 70%	88%	88%	N/A	Pass

**Summary of observations:** All test and control sperm was prepared from the same donor and incubated in the same incubator at 32°C and 5% CO<sub>2</sub>. The control sperm had an 88% forward progressive motility at 24-hours. The test article sperm had an 85% forward progressive motility at 24-hours.

Signature  
Study Director

03-24-2017  
Date

Signature  
Quality Reviewer

03-24-2017  
Date