



SparMED Aps  
Ryttermarken 2  
3520 Farum  
Denmark



ELI Accession Number: SPAR-9696-1018

Date of completion: 11-03-2018

Lot number: 07965

Reference number: OOPW-AT10

Description of test article(s): Oosafe® Andrology Tube

Assay system requested by customer: 1mL of the culture medium was placed in the test article and overlaid with oil. 21 one cell mouse embryos were placed in the test article and cultured for 96-hours.

Control assay method and results: 15 one cell (B6C3F1 X B6D2F1) embryos were cultured in 0.5mL drops of culture medium overlaid with oil in a 4-well dish:

15 / 15 (100 %)

1-cell to 2-cell within 24 hr

15 / 15 (100 %)

1-cell to expanded blastocyst within 96 hr

For a valid assay, *Embryotech™* requires at least 70% of one cell stage control embryos to develop to expanded blastocyst within 96-hours.

Test assay method and results: 21 one cell (B6C3F1 X B6D2F1) embryos were cultured in the test article using culture medium:

21 / 21 (100 %)


1-cell to 2-cell within 24 hr

21 / 21 (100 %)


1-cell to expanded blastocyst within 96 hr

Summary of observations: All test and control embryos were selected randomly from a common pool of freshly collected embryos and were cultured in the same incubator at 37°C and 5.0% CO<sub>2</sub>. 100 percent of the control embryos developed to the expanded blastocyst stage within 96-hours. 100 percent of the embryos cultured in the test article developed to the expanded blastocyst stage within 96-hours.

  
\_\_\_\_\_  
Signature  
Study Director

  
\_\_\_\_\_  
Date

  
\_\_\_\_\_  
Signature  
Quality Reviewer

  
\_\_\_\_\_  
Date