

Doc. No.: QC_4.14.2. Rev. Date: 01.04.2019



SparMED ApS

VAT no.: DK 30 89 85 75

Ryttermarken 2, 3520, Farum, Denmark

Phone: +45 39 40 25 03

e-mail: info@sparmed.dk - www.sparmed.dk

Certificate of Analysis

Product ID: Oosafe® Plasticware: OOSAFE® Center well dish with two compartments

Product code: OOPW-CW05

LOT No.: 07995

Expiry date: 2024-04

Storage Conditions: 20-30°C, dry room, no exposure to sunlight

Quality Assurance:

Analysis	Reference No.	Results	Completion date
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-1810-0519	Passed	26-05-2019
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S3863-0519SPAR	Passed	24-05-2019
Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria <0.03 EU/device.	E9325-0519SPAR	Passed	22-05-2019
Sterilization by gamma irradiation. Delivered irradiation dose: 8.6 kGy-9.5 kGy. Specified irradiation dose: 8.0 kGy-10.0 kGy.	N/A	Passed	N/A

Quality Control according to ISO 13485:2016

GOosafe with SparMED!

Qi Wei,

Quality Control Manager

SparMED ApS

Date: 03-06-2019

