



Doc. No.: QC\_4.14.2. Rev. Date: 01.04.2019



**sparMED**  
GOosafe in IVF Laboratories

**SparMED ApS**

VAT no.: DK 30 89 85 75

Ryttermarken 2, 3520, Farum, Denmark

Phone: +45 39 40 25 03

e-mail: [info@sparmed.dk](mailto:info@sparmed.dk) - [www.sparmed.dk](http://www.sparmed.dk)

## Certificate of Analysis

**Product ID:** Oosafe® OPU Tube, 14 ml

**Product code:** OOPW-OT10

**LOT No.:** 08005

**Expiry date:** 2024-07

**Storage Conditions:** 20-30°C, dry room, no exposure to sunlight

**Quality Assurance:**

Analysis	Reference No.	Results	Completion date
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-2389-0819	Passed	27-08-2019
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S4014-0819SPAR	Passed	27-08-2019
Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria <0.03 EU/device.	E9545-0819SPAR	Passed	24-08-2019
Sterilization by gamma irradiation. Delivered irradiation dose: 8.6 kGy-9.5 kGy. Specified irradiation dose: 8.0 kGy-10.0 kGy.	N/A	Passed	N/A

Quality Control according to ISO 13485:2016

**GOosafe with SparMED!**

Inga Sileikaite,  
Quality Control Department  
SparMED ApS

**Date: 28-08-2019**

