

Certificate of Analysis

Product ID: OOSAFE® 100 MM DISH

Product code: OOPW-HD10

LOT No.: 08023

Expiry date: 2024-10

Storage Conditions: 20-30°C, dry room, no exposure to sunlight

Quality Assurance:

Analysis	Reference No.	Results	Completion date
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-3264-0120	Passed	21-01-2020
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S4251-0120SPAR	Passed	22-01-2020
Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria <0.03 EU/device.	E9911-0120SPAR	Passed	17-01-2020
Sterilization by gamma irradiation. Delivered irradiation dose: 8.6 kGy-9.5 kGy. Specified irradiation dose: 8.0 kGy-10.0 kGy.	NA	Passed	NA

Quality Control according to ISO 13485:2016

GOosafe with SparMED!

Inga Sileikaite,
Quality Control Department
SparMED ApS

Date: 23-01-2020

