

Certificate of Analysis

Product ID: Oosafe® OPU Tube

Product code: OOPW-OT10

LOT No.: 08034

Expiry date: 2024-11

Storage Conditions: 20-30°C, dry room, no exposure to sunlight

Quality Assurance:

Analysis	Reference No.	Results	Completion date
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-3231-0120	Passed	17-01-2020
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S4242-0120SPAR	Passed	15-01-2020
Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria <0.03 EU/device.	E9894-0120SPAR	Passed	17-01-2020
Sterilization by gamma irradiation. Delivered irradiation dose: 8.6 kGy-9.5 kGy. Specified irradiation dose: 8.0 kGy-10.0 kGy.	NA	Passed	NA

Quality Control according to ISO 13485:2016

GOosafe with SparMED!

Inga Sileikaite,
Quality Control Department
SparMED ApS

Date: 20-01-2020

