

Certificate of Analysis

Product ID: Oosafe® 4 Well Dish, Non-Treated Surface

Product code: OOPW-FW04

LOT No.: 08052

Expiry date: 2025-04

Storage Conditions: 20-30°C, dry room, no exposure to sunlight

Quality Assurance:

Analysis	Reference No.	Results	Completion date
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-4248-0720	Passed	12-07-2020
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S4530-0720SPAR	Passed	10-07-2020
Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria <0.03 EU/device.	E0332-0720SPAR	Passed	08-07-2020
Sterilization by gamma irradiation. Delivered irradiation dose: 8.6 kGy-9.5 kGy. Specified irradiation dose: 8.0 kGy-10.0 kGy.	N/A	Passed	N/A

Quality Control according to ISO 13485:2016

GOosafe with SparMED!

Florencia Garcia
Quality Control Department
SparMED ApS
Date: 03-08-2020


