

Certificate of Analysis

Product ID: Oosafe® 35 mm Dish, High Wall

Product code: OOPW-TF03

LOT No.: 08080

Expiry date: 2025-10

Storage Conditions: 20-30°C, dry room, no exposure to sunlight

Quality Assurance:

Analysis	Reference No.	Results	Completion date
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-5255-0221	Passed	13-02-2021
Proved stable human sperm motility: $\geq 75\%$ sperm motility after 24 hours proven.	S4853-0221SPAR	Passed	10-02-2021
Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria < 0.03 EU/device.	E0804-0221SPAR	Passed	09-02-2021
Sterilization by gamma irradiation. Delivered irradiation dose: 8.6 kGy-9.5 kGy. Specified irradiation dose: 8.0 kGy-10.0 kGy.	N/A	Passed	N/A

Quality Control according to ISO 13485:2016

GOosafe with SparMED!

Florencia Garcia
Quality Control Department
SparMED ApS
Date: 17-02-2021

