

Certificate of Analysis

Product ID: OOSAFE® CENTER WELL DISH

Product code: OOPW-CW08

LOT No.: 08081

Expiry date: 2025-10

Storage Conditions: 20-30°C, dry room, no exposure to sunlight

Quality Assurance:

Analysis	Reference No.	Results	Completion date
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-7894-0322	Passed	21-03-2022
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S5725-0322SPAR	Passed	22-03-2022
Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria <0.03 EU/device.	E2095-0322SPAR	Passed	21-03-2022
Sterilization by gamma irradiation. Delivered irradiation dose: 8.6 kGy-9.5 kGy. Specified irradiation dose: 8.0 kGy-10.0 kGy.	N/A	Passed	N/A

Quality Control according to ISO 13485:2016

GOosafe with SparMED!

Andrius Daugintis
Quality Control Department
SparMED ApS
Date: 06-04-2022

