

Certificate of Analysis

Product ID: Oosafe® OPU Tube

Product code: OOPW-OT10

LOT No.: 08084

Expiry date: 2025-10

Storage Conditions: 20-30°C, dry room, no exposure to sunlight

Quality Assurance:

Analysis	Reference No.	Results	Completion date
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-5026-1220	Passed	26-12-2020
Proved stable human sperm motility: $\geq 75\%$ sperm motility after 24 hours proven.	S4763-1220SPAR	Passed	22-12-2020
Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria < 0.03 EU/device.	E0661-1220SPAR	Passed	21-12-2020
Sterilization by gamma irradiation. Delivered irradiation dose: 8.6 kGy-9.5 kGy. Specified irradiation dose: 8.0 kGy-10.0 kGy.	N/A	Passed	N/A

Quality Control according to ISO 13485:2016

GOosafe with SparMED!

Florencia Garcia
Quality Control Department
SparMED ApS
Date: 08-01-2021

