



Doc. No.: QC_4.14.2. Rev. Date: 01.04.2019



SparMED ApS

VAT no.: DK 30 89 85 75

Ryttermarken 2, 3520, Farum, Denmark

Phone: +45 39 40 25 03

e-mail: info@sparmed.dk - www.sparmed.dk

Certificate of Analysis

Product ID: Oosafe® Andrology Tube, 5 ml

Product code: OOPW-AT10

LOT No.: 08130

Expiry date: 2026-05

Storage Conditions: 20-30°C, dry room, no exposure to sunlight

Quality Assurance:

Analysis	Reference No.	Results	Completion date
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-6168-0721	Passed	24-07-2021
Proved stable human sperm motility: $\geq 75\%$ sperm motility after 24 hours proven.	S5141-0721SPAR	Passed	21-07-2021
Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria < 0.03 EU/device.	E1298-0721SPAR	Passed	20-07-2021
Sterilization by gamma irradiation. Delivered irradiation dose: 8.6 kGy-9.5 kGy. Specified irradiation dose: 8.0 kGy-10.0 kGy.	N/A	Passed	N/A

Quality Control according to ISO 13485:2016

GOosafe with SparMED!

Zahraa Baker
Quality Control Department
SparMED ApS
Date: 28-07-2021

