



Doc. No.: QC_4.14.2. Rev. Date: 01.04.2019



sparMED
GOosafe in IVF Laboratories

SparMED ApS

VAT no.: DK 30 89 85 75

Ryttermarken 2, 3520, Farum, Denmark

Phone: +45 39 40 25 03

e-mail: info@sparmed.dk - www.sparmed.dk

Certificate of Analysis

Product ID: OOSAFE® 6 Well dish with straw holder

Product code: OOPW-SW02

LOT No.:08131

Expiry date: 2026-05

Storage Conditions: 20-30°C, dry room, no exposure to sunlight

Quality Assurance:

Analysis	Reference No.	Results	Completion date
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-6605(R1)-0821	Passed	12-09-2021
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S5287-0821SPAR	Passed	03-09-2021
Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria <0.03 EU/device.	E1451-0821SPAR	Passed	01-09-2021
Sterilization by gamma irradiation. Delivered irradiation dose: 8.6 kGy-9.5 kGy. Specified irradiation dose: 8.0 kGy-10.0 kGy.	N/A	Passed	N/A

Quality Control according to ISO 13485:2016

GOosafe with SparMED!

Andrius Daugintis
Quality Control Department
SparMED ApS
Date: 01-10-2021

