

## Certificate of Analysis

**Product ID: Oosafe® 4 Well Dish, Non-Treated Surface****Product code: OOPW-FW04****LOT No.: 08188****Expiry date: 2026-11****Storage Conditions: 20-30°C, dry room, no exposure to sunlight****Quality Assurance:**

Analysis	Reference No.	Results	Completion date
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-7256-1221	Passed	14-12-2021
Proved stable human sperm motility: $\geq 75\%$ sperm motility after 24 hours proven.	S5527-1221SPAR	Passed	16-12-2021
Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria $< 0.03$ EU/device.	E1795-1221SPAR	Passed	11-12-2021
Sterilization by gamma irradiation. Delivered irradiation dose: 8.6 kGy-9.5 kGy. Specified irradiation dose: 8.0 kGy-10.0 kGy.	N/A	Passed	N/A

Quality Control according to ISO 13485:2016

**GOosafe with SparMED!**

Zahraa Baker  
Quality Control Department  
SparMED ApS  
Date: 05-01-2022

