



Doc. No.: QC_4.14.2. Rev. Date: 01.04.2019



SparMED
GOosafe in IVF Laboratories

SparMED ApS

VAT no.: DK 30 89 85 75

Ryttermarken 2, 3520, Farum, Denmark

Phone: +45 39 40 25 03

e-mail: info@sparmed.dk - www.sparmed.dk

Certificate of Analysis

Product ID: OOSAFE® Andrology Tube, 5 ml

Product code: OOPW-AT10

LOT No.: 08214

Expiry date: 2027-01

Storage Conditions: 20-30°C, dry room, no exposure to sunlight

Quality Assurance:

Analysis	Reference No.	Results	Completion date
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-7568-0122	Passed	04-02-2022
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S5621-0122SPAR	Passed	10-02-2022
Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria <0.03 EU/device.	E1940-0122SPAR	Passed	01-02-2022
Sterilization by gamma irradiation. Delivered irradiation dose: 8.6 kGy-9.5 kGy. Specified irradiation dose: 8.0 kGy-10.0 kGy.	N/A	Passed	N/A

Quality Control according to ISO 13485:2016

GOosafe with SparMED!

Zahraa Baker
Quality Control Department
SparMED ApS
Date: 14-02-2022

