



Doc. No.: QC\_4.14.2. Rev. Date: 01.04.2019



**SparMED ApS**

VAT no.: DK 30 89 85 75

Ryttermarken 2, 3520, Farum, Denmark

Phone: +45 39 40 25 03

e-mail: [info@sparmed.dk](mailto:info@sparmed.dk) - [www.sparmed.dk](http://www.sparmed.dk)

## Certificate of Analysis

**Product ID: Oosafe® 4 Well Dish, Non-Treated Surface**

**Product code: OOPW-FW04**

**LOT No.: 08232**

**Expiry date: 2027-02**

**Storage Conditions: 20-30°C, dry room, no exposure to sunlight**

**Quality Assurance:**

Analysis	Reference No.	Results	Completion date
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-8092-0422	Passed	18-04-2022
Proved stable human sperm motility: ≥75% sperm motility after 24 hours proven.	S5775-0422SPAR	Passed	15-04-2022
Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria <0.03 EU/device.	E2179-0422SPAR	Passed	15-04-2022
Sterilization by gamma irradiation. Delivered irradiation dose: 8.6 kGy-9.5 kGy. Specified irradiation dose: 8.0 kGy-10.0 kGy.	N/A	Passed	N/A

Quality Control according to ISO 13485:2016

### GOosafe with SparMED!

Zahraa Baker  
Quality Control Department  
SparMED ApS  
Date: 21-04-2022

