



Doc. No.: QC_4.14.2. Rev. Date: 01.04.2019

SparMED ApS



sparMED

GOosafe in IVF Laboratories

VAT no.: DK 30 89 85 75

Ryttermarken 2, 3520, Farum, Denmark

Phone: +45 39 40 25 03

e-mail: info@sparmed.dk - www.sparmed.dk

Certificate of Analysis

Product ID: Oosafe® Snap-Cap Tube 14 ml

Product code: OOPW-OT10

LOT No.: 08241

Expiry date: 2027-04

Storage Conditions: 20-30°C, dry room, no exposure to sunlight

Quality Assurance:

Analysis	Reference No.	Results	Completion date
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-8324-0522	Passed	30-05-2022
Proved stable human sperm motility: $\geq 75\%$ sperm motility after 24 hours proven.	S5850-0522SPAR	Passed	27-05-2022
Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria < 0.03 EU/device.	E2287-0522SPAR	Passed	26-05-2022
Sterilization by gamma irradiation. Delivered irradiation dose: 8.6 kGy-9.5 kGy. Specified irradiation dose: 8.0 kGy-10.0 kGy.	N/A	Passed	N/A

Quality Control according to ISO 13485:2016

GOosafe with SparMED!

Zahraa Baker
Quality Control Department
SparMED ApS
Date: 09-06-2022

