

SparMED ApS

VAT no.: DK 30 89 85 75

Ryttermarken 2, 3520, Farum, Denmark

Phone: +45 39 40 25 03

e-mail: info@sparmed.dk - www.sparmed.dk

Certificate of Analysis

Product ID: OOSAFE® 4 Well Dish, Non-Treated Surface**Product code: OOPW-FW04****LOT No.: 08242****Expiry date: 2027-04****Storage Conditions: 20-30°C, dry room, no exposure to sunlight****Quality Assurance:**

Analysis	Reference No.	Results	Completion date
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-8259-0522	Passed	20-05-2022
Proved stable human sperm motility: $\geq 75\%$ sperm motility after 24 hours proven.	S5824-0522SPAR	Passed	18-05-2022
Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria < 0.03 EU/device.	E2253-0522SPAR	Passed	17-05-2022
Sterilization by gamma irradiation. Delivered irradiation dose: 8.6 kGy-9.5 kGy. Specified irradiation dose: 8.0 kGy-10.0 kGy.	N/A	Passed	N/A

Quality Control according to ISO 13485:2016

GOosafe with SparMED!Zahraa Baker
Quality Control Department
SparMED ApS
Date: 08-07-2022