

Certificate of Analysis

Product ID: OOSAFE® SNAP-CAP TUBE

Product code: OOPW-OT10

LOT No.: 08251

Expiry date: 2027-05

Storage Conditions: 15-30°C, dry room, no exposure to sunlight

Quality Assurance:

Analysis	Reference No.	Results	Completion date
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-8462-0622	Passed	21-06-2022
Proven stable human sperm motility: $\geq 75\%$ sperm motility after 24 hours proven.	S5901-0622SPAR	Passed	18-06-2022
Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria < 0.03 EU/device.	E2352-0622SPAR	Passed	17-06-2022
Sterilization by gamma irradiation. Specified irradiation dose: 10.0 kGy-12.0 kGy.	N/A	Passed	N/A

Quality Control according to ISO 13485:2016

GOosafe with SparMED!

Qi Wei

Quality Control Department

SparMED ApS

Date: 22/06/2022

