

Certificate of Analysis

Product ID: Oosafe® Snap-Cap tube**Product code:** OOPW-OT10**LOT:** 08259**Expiry date:** 2027-07**Storage Conditions:** 15-30°C, dry room, no exposure to sunlight**Quality Assurance:**

Analysis	Reference No.	Results	Completion date
Proven non-embryotoxic by Mouse Embryo Assay Test. Over 80% embryo development to the expanded blastocyst stage within 96 hours.	SPAR-9289-0822	Passed	14-08-2022
Proven stable human sperm motility: $\geq 75\%$ sperm motility after 24 hours proven.	S6012-0822SPAR	Passed	11-08-2022
Proven non-toxic by Limulus Amebocyte Lysate (LAL) test. Pass criteria < 0.03 EU/device.	E2503-0822SPAR	Passed	11-08-2022
Sterilized by E-Beam sterilization. Specified sterilization dose: 10.0 kGy-12.0 kGy.	N/A	Passed	N/A

Quality Control according to ISO 13485:2016

GOosafe with SparMED!

Andrius Daugintis,

Quality Control

SparMED ApS

Date: 02/09/2022