

# Certificate of Analysis

**Product name: Oosafe® 60 mm Dish, Non-Treated**

**Product type: Oosafe® Plasticware IVD**

**Product code: OOPW-ST06**

**Batch/LOT No.: 08972**

**Product expiry date: 2028-10**

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	CONCLUSION	DOCUMENT NO.
Mouse Embryo Assay Test (MEA)	Over 80% embryo development to the expanded blastocyst stage within 96 hours	Passed	Non-embryotoxic	SPAR-3511-0124
Human Sperm Survival Assay (HSSA)	≥75% sperm motility after 24 hours proven	Passed	Stable human sperm motility	S7027-0124SPAR
Limulus Amoebocyte Lysate (LAL)	<0.03 EU/device	Passed	Do not contain endotoxin	E3719-0124SPAR

Quality Management system according to ISO 13485:2016

**Sterilized by irradiation according to ISO 11137-1**

**Sterilization method: E-Beam**

**Sterility assurance level: 10<sup>-6</sup>**

**GOosafe with SparMED!**

Manuela Jagieniak




Quality Control Department

SparMED ApS

Issue date: 01.02.2024



**Sparmed ApS**

Ryttermarken 2, 3520 Farum, DK  +45 39 40 25 03  info@sparmed.dk  www.sparmed.dk  
Corp. VAT no.: DK 30 89 85 75