

Certificate of Analysis

Product name: Center Well Dish, Non-Treated

Product type: Oosafe® Plasticware IVD

Product code: OOPW-CW08

Batch/LOT No.: 08994

Product expiry date: 2029-03

TESTING METHOD	ACCEPTANCE CRITERIA	RESULTS	CONCLUSION	DOCUMENT NO.
Mouse Embryo Assay Test (MEA)	Over 80% embryo development to the expanded blastocyst stage within 96 hours	Passed	Non-embryotoxic	SPAR-3881-0324
Human Sperm Survival Assay (HSSA)	≥75% sperm motility after 24 hours proven	Passed	Stable human sperm motility	S7155-0324SPAR
Limulus Amoebocyte Lysate (LAL)	<0.03 EU/device	Passed	Do not contain endotoxin	E3859-0324SPAR

Quality Management system according to ISO 13485:2016

Sterilized by irradiation according to ISO 11137-1

Sterilization method: E-Beam

Sterility assurance level: 10⁻⁶

GOosafe with SparMED!

Manuela Jagieniak

Quality Control Department

SparMED ApS

Issue date: 11.04.2024



Sparmed ApS

Ryttermarken 2, 3520 Farum, DK ● +45 39 40 25 03 ● info@sparmed.dk ● www.sparmed.dk
Corp. VAT no.: DK 30 89 85 75